



POLISH CENTRE FOR TESTING AND CERTIFICATION

02-699 Warszawa, ul. Kłobucka 23A

Mechanical Laboratory

tel.: (+48 22) 46 45 594; fax: (+48 22) 46 45 563

e-mail: labmech@pcbc.gov.pl

Date 23.12.2011

TEST REPORT NO. BR -243/L-349/2011

Page 1 of 4

Subject of testing:	<i>Assistive products for walking manipulated by one arm Elbow crutch</i>	Classification according to PN-EN ISO 9999:2007 : 12 03 06
Type / Model:	315021	Factory ref. no.: --
Manufacturer:	MOBILEX A/S, Nørskovvej 1, DK - 8660 Skanderborg	Number of specimens: 2
Applicant:	A-Net s.c., ul. Łaskowice 174, 93-469 Łódź	
Kind of testing	<i>Mechanical testing for conformity with</i> PN-EN 11334-1:2007 - <i>clause 4,6; methods - clause. 5</i>	

Test started: 5.12.2011

Test finished: 23.12.2011

Performed by:

Mirosław Szymański

Checked by:

Ireneusz Czerwiński

Approved by:

KIEROWNIK
LABORATORIUM MECHANICZNEGO

mgr inż. Andrzej Tkaczyk

Special comments / enclosures:

1) labels, service manual

Copyright © 2005 by Polish Centre for Testing and Certification (applicable to report form)

Test results refer only to tested units.

This test report shall be neither copied differently as in the whole nor be published without written consent of the Laboratory.

Constructional, material or technological modifications made in product require re-assessment of product conformity with requirements of above mentioned standards



CHARACTERISTIC OF ELBOW CRUTCH

Name : *Elbow crutch*

Maximum permissible user mass: *150 kg*

Mass of crutch: *0,55 kg*

Description		Comments
Dimensions of crutch:	Overall height of crutch:	min. 999 mm
		max. 1230 mm
	Leg section length (l):	min. 778 mm
		max. 1006 mm
		number of fixing positions 10
	Arm section length (a):	min. 210 mm
		max. --
		number of fixing positions --
	Handgrip length (h):	100 mm
	Handgrip width:	35,6 mm
	Cuff internal width (y):	80 mm
	Cuff internal depth (x):	85 mm
	Cuff internal height (z):	<i>cuff and arm section are one element of the crutch</i>
	Support angle (α):	22°
Grip angle (β):	105°	<i>According to figure 1</i>
Tip diameter:	Ø 39 mm	
Material of	Leg section :	<i>Tubes – aluminum alloy Ø21,95 mm, Ø18,95 mm</i>
	Handgrip :	<i>Plastic Gray colour</i>
	Arm section:	<i>Plastic Gray colour</i>
	Cuff:	<i>Plastic Gray colour</i>
	Tip:	<i>Rubber Gray colour</i>
Reflective elements	back	<i>Not included</i>
	front	<i>Included</i>

PHOTO OF ELBOW CRUTCH:




MOBILEX A/S
Nørskovvej 1
DK - 8660 Skanderborg
Tel: +45 87 93 22 20
www.mobilex-care.com

Elbow crutch



CE Produced Serial no.



RESULT OF TESTS ACCORDING TO PN-EN ISO 11334-1:2007

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters	Real value	Test result	Comments	
1	2	3	4	5	6	
	Measur.	Dimensions of crutch	Overall height of crutch (min.)	999 mm	N/R.	
	Measur.		Overall height of crutch (max.)	1230 mm	N/R.	
	Measur.		Leg section length ($l_{min.}$)	778 mm	N/R.	
	Measur.		Leg section length ($l_{max.}$)	1006 mm	N/R.	
	Measur.		Arm section length ($a_{min.}$)	210 mm	N/R.	
	Measur.		Arm section length ($a_{max.}$)			
	Measur.		Handgrip length (h)	100 mm	N/R.	
	Measur.		Handgrip width (S)	35,6mm	N/R.	
	Measur.		Cuff internal depth (x)	85 mm	N/R.	
	Measur.		Cuff internal with (y)	80 mm	N/R.	
	Measur.		Cuff internal height (z)	--	N/A	cuff and arm section are one element of the crutch
	Measur.		Width of opening in cuff	--	N/A	
	Measur.		Support angle α	22°	N/R.	
	Measur.		Grip angle β	105°	N/R.	
	Measur.	Tip diameter	39 mm	N/R.		
	Measur.	Mass of crutch	0,55kg	0,55kg		
4.1	V/I	Cuff	Securely fixing of cuff	Conf.	Pos.	
	Measur.		Internal cuff dimensions	Conf.	Pos.	
	V/I		Position of internal cuff surface towards its support line	Conf.	Pos.	
	Measur.		Relation of internal cuff depth to its internal width	Conf. (y)- Cuff width 80mm (x)- Cuff depth 85mm	Pos.	Required internal depth of cuff more than half of its internal width
	Measur.		Cuff internal height (z)	N/A	Pos.	Required internal height of cuff ≥ 40 mm Cuff and arm section are one element
	5.4		Withdrawal test	N/A	Pos.	Required withdrawal force not more than 120N
4.2	V/I	Handgrip	Securely fixing of handgrip	Conf.	Pos.	
	Measur.		Handgrip resistance to sliding of hand when crutch is in use	Conf.	Pos.	
	V/I		Handgrip width :	35,6 mm Conf.	Pos.	Handgrip width required ≥ 25 mm $i \leq 50$ mm
	Measur.		Ease of clearing	Conf.	Pos.	
4.3	5.3	Leg section and tip	Construction, tip characteristics	Conf.	Pos.	
	Measur.		Tip diameter	$\emptyset 39$ mm Conf.	Pos.	Tip diameter required ≥ 35 mm
4.4	V/I	Adjusting devices	Fastness to loosening of height adjustment elements	Conf.	Pos.	
	Measur.		Maximum extension of the height adjustment, marked on the adjusting members	Conf.	Pos.	
	V/I		Possibility of adjustment elements operation without use of tools	Conf.	Pos.	
4.5	V/I	Materials	--	N/T	Manufacturer's declaration	

Requirements according to clause	Test method according to clause	Checked characteristics/assemblies/parameters		Real value	Test result	Comments
1	2	3		4	5	6
4.6	5.5	Mechanical durability	Crutch element separation test (including handgrip)	Conf.	Pos.	loading 500N (handgrip 750N)
	5.6		Static loading test	Conf.	Pos.	loading 1500N
	5.7		Fatigue strength of crutch	1 000 000 cycles	Pos.	loading 825N
	5.8		Low temperature falling test	Conf.	Pos.	
PN-EN 1041	V/I	Information supplied by the manufacturer		Conf.	Pos.	
6.2 6.3	V/I	Marking and labelling		Conf.	Pos.	
Test conditions:		Temperature of air		21°C	Temperature required 21°C +/-2°C	
		Relative humidity of air:		41%	N/R.	

Comments:

All tests were performed at maximum height of elbow crutches

Two elbow crutches of the same type and kind were tested: one was put to fatigue test and separation test, the other was put to static loading test and low temperature falling test.

Sequence of tests: measurements, separation test, static loading test, fatigue test, stability test, low temperature falling test.

During visual inspection before testing any visible defects that could have influence on test results were not stated.

Pos. – positive; Neg – negative; N/T – not tested; N/A – not applicable; N/R – not required, N/O – not occurred, V/I.- visual inspection, Conf.- conformed.

CONCLUSIONS:

Test object **conforms** with requirements of PN-EN ISO 11334-1:2007 - „Assistive products for walking manipulated by one arm. Elbow crutches. Requirements and test methods” – clause 4, 6 (methods – clause. 5) excluding testing of biological conformity of material with human body according to PN-EN ISO 10993-1:2010.

---- END ----

INFORMATION SUPPLIED BY THE MANUFACTURER

Name of product: *Elbow crutch*Applicant: *A-Net s.c.**93-469 Łódź, ul. Łaskowice 174*

Requirement according to PN-EN ISO 11334-1:2007		Included
6.2 Information of product		
6.2a	Maximum permissible user mass	YES
6.2b	Manufacturer's name or trade name and address	YES
6.2c	Manufacturer's model identification name and/or number	YES
6.2d	Month and year of manufacture	YES
6.2e	Maximum extension of the height adjustment, marked on the adjusting members	YES
Information in documentation, on tag or on product		
6.3a	Maximum arm section length a_{max}	YES
6.3b	Minimum arm section length a_{min}	YES
6.3c	Maximum leg section length l_{max}	YES
6.3d	Minimum leg section length. l_{min}	YES
6.3e	Support angle α	YES
6.3f	Cleaning instructions, including a description of the method and suitable cleaning agents and any precautions needed to avoid corrosion and/or ageing of the materials used in construction of the elbow crutch	YES
6.3g	Instructions for assembly, adjustment of all kinds, folding and unfolding, if applicable	YES
6.3h	Warnings and advice about precautions relating to safe distances between moving and stationary parts, if applicable (see EN 12182:2005, Clauses 12 and 13, for guidance)	N/A
CE Marking		YES

N/A- not applicable